


# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130295
Name	Charles Foster	ID#	7734
Date	5/5/11		
A	Agency	SCDL	Phone # 9072695760
	Instrument Location	SCDL	
Alco S/N		97073U	Target Value 0.080
High Pressure		500	
B	Alco Test Values	0.081	0.081
	1 <sup>st</sup> Alco		2 <sup>nd</sup> Alco
Signature			
(OVER)			


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of June, 2010.

 (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130295

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130295

MAY 05, 2011

OPERATOR'S NAME:  
FOSTER/CHARLES  
OPERATOR'S NUMBER: 7734  
SUBJECT'S LAST NAME:  
UOC

SUBJECT'S FIRST NAME/MI :  
UOC

O.L. #: UOC  
DEPT/AGENCY: UOC  
CASE/REPORT: UOC  
TEST TYPE: T  
ALCO TARGET VALUE: 0.080  
ALCO S/N: 970731

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.73 in  
ALCO TARGET .079 12:59  
BLANK TEST .000 13:00  
INTERNAL STANDARD VERIFIED 13:00  
ALCO TO 29.72 in .081 13:00  
BLANK TEST .000 13:01  
SUBJECT SAMPLE .000 13:01  
BLANK TEST .000 13:02  
ALCO TO 29.73 in .081 13:02  
BLANK TEST .000 13:03

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130295

MAY 05, 2011  
TIME 13:06

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/02/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 43c  
BAROMETER: 29.73 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

11432'()++-./0123456789:;<=>@ABCDEF  
GHIJKLMNOPQRSTUVWXYZ[\]^\_`abcde  
fghijklmnopqrstuvwxyz{|}~